

REMARKS

Applicants appreciate the notification in the Office Action of allowable subject matter, i.e. that claims 1-9 and 20 are allowed.

Claims 10 and 14 have been amended. No new matter has been added by virtue of the amendments. For instance, support for the amendments appears in the original claims.

Claims 10-19 were rejected under 35 U.S.C. 112 second paragraph. The language of "one can re-insert" was objected to.

While Applicants fully disagree with the rejection, the language noted in the Office Action also has been amended herein. It is thus believed the rejection has been obviated.

Reconsideration and withdrawal of the rejection are requested.

Claims 10-14 were rejected under 35 U.S.C. 103 over Yamada in view of Sahatijian and optionally further in view of Krasner. The rejection is traversed.

Applicants' independent claim 10 (the only independent claim rejected over Yamada/Sahatijian/Krasner) calls for a method for implanting an intraocular lens in an eye wherein after a movable member is removed from the eye and an intraocular lens remains in the eye, an injection device is inserted into the intraocular lens to adjust the amount of optical medium in the intraocular lens


Yamada does not describe or suggest a device or method as Applicants claim, including wherein after a lens is inflated and left in the eye, at a later date one can re-insert an injection device into the intraocular lens and adjust the amount of optical medium in the intraocular lens.

The secondary citations of Sahatjian and Krasner either alone or in combination do not remedy such deficiencies of Yamada.

In view thereof, reconsideration and withdrawal of the rejection are requested.

It is believed the application is in condition for immediate allowance, which action is earnestly solicited.

Respectfully submitted,

A handwritten signature in black ink, appearing to read 'P. Corless', written over a horizontal line.

Peter F. Corless (Reg. No. 33,860)
EDWARDS & ANGELL, LLP
P.O. Box 55874
Boston, Massachusetts 02205
Tel. No. (617) 517-5512